BJ-3 DISINTEGRATION TESTER



BJ-3 DISINTEGRATION TESTER

BJ series of instruments are used for detecting disintegration of solid in prescriptive conditions.

Applicable standards:

National standard (Chinese Codex) /2015 Corporate standard (Dissolution tester)Q/12XQ0188-2010

Features:

There are eighteen tubes in three basket assembly.

Each test station is driven independently. Auto stop at timing point.

The bath liquid can achieve an even temperature.

A MPU is used to automatically control the temperature, time and frequency of nacelle back and forth.

Timing at will. The preset and real time data can be displayed alternately in time.

Automation: auto-test, auto-diagnose, auto-alarm.

Specifications:

Basket assembly: 3

Nacelle back and forth frequency: $(30\sim32)$ /min

Nacelle back and forth range: (55 ± 1) mm Automatic time range: $(5\sim120)$ min

Accuracy of controlling temperature: $(37\pm0.5)^{\circ}$ C Power: 220V/50Hz/600W or 110V/60Hz/600W

Dimension: (540×300×420)mm3